TRC
Results you can rely on

6312 NW 18th Drive
Suite 100
Gainesville, FL 32653

352.378.0332 PHONE
352.378.0354 FAX

www.TRCsolutions.com

December 6, 2016

Ms. Lori Simmons
Arkansas Department of Health
4815 West Markham Street
Little Rock, Arkansas 72205
Via email Lori.Simmons@arkansas.gov

Re: Georgia-Pacific, Crossett Mill - Biweekly Air Monitoring Report for Hydrogen Sulfide

Dear Ms. Simmons,

Following is the biweekly data summary for the Georgia-Pacific (GP) hydrogen sulfide (H_2S) and meteorological monitoring program, at the GP Crossett mill, covering the calendar period of November 16th through November 29th.

Summary of Results

Included in this report are three plots presenting H_2S concentrations calculated with varied rolling average periods (30-minute, 8-hour, and 24-hour).

Also included in this report is a summary of results from the daily 1-point QC checks performed during this biweekly period. The QAPP establishes goals for precision and bias as a coefficient of variation (CV) <10% and ± 10%, respectively. Precision and bias are calculated in accordance with 40 CFR Part 58 Appendix A, Section 4.1.

There were two occurrences of data loss during this monitoring period, in addition to those resulting from automated daily 1-point QC and weekly calibration checks. On the afternoon of November 17th, TRC performed a manual adjustment of zero and span calibration concentrations. As a result, approximately two hours of H_2S data was lost on the 17th. On the afternoon of November 29th, a complete calibration was performed, resulting in approximately one and a half hours of data loss. Results for all automated daily 1-point QC checks fall within the acceptable range, indicating the H_2S monitor was operating in accordance with the QAPP.

Fourteen-day time series plots for all recorded meteorological (met) parameters are presented in the final table. All met parameters have 100% data capture for this report period.

Please feel free to contact me if you have any questions or need any additional data.

9846255

TRC

Sincerely,

Jonathan Bowser
Manager, Air Quality and Meteorological Monitoring

Air Measurements – Gainesville Office
6312 NW 18th Drive, Suite 100
Gainesville, Florida 32653
(352) 260-1162
Email: jbowser@trcsolutions.com

CC: Becky Keough, ADEQ Director via email: keogh@adeq.state.ar.us
Kara Allen, Environmental Engineer, USEPA Region 6 via email Allen.Kara@epa.gov


H2S 30 Min Rolling Avg
Georgia Pacific Crossett, AR
H2S (ppb)
100.0
90.0
80.0
70.0
60.0
50.0
40.0
30.0
20.0
10.0
0.0
11/16
11/17
11/18
11/19
11/20
11/21
11/22
11/23
11/24
11/25
11/26
11/27
11/28
11/29


H2S 8 Hr Rolling Avg
Georgia Pacific Crossett, AR
H2S (ppb)
100.0
90.0
80.0
70.0
60.0
50.0
40.0
30.0
20.0
10.0
0.0
11/16
11/17
11/18
11/19
11/20
11/21
11/22
11/23
11/24
11/25
11/26
11/27
11/28
11/29


H2S 24 Hr Rolling Avg
Georgia Pacific Crossett, AR
H2S (ppb)
100.0
90.0
80.0
70.0
60.0
50.0
40.0
30.0
20.0
10.0
0.0
11/16
11/17
11/18
11/19
11/20
11/21
11/22
11/23
11/24
11/25
11/26
11/27
11/28
11/29

H_2S Assessment

GP - Crossett, AR			Compound of Interest: H_2S					CV_{ub} (%)		Bias (%)	
Date	Meas Val (Y)	Audit Val (X)	d (Eqn. 1)	25th Percentile	d^2	|d|	$\|d\|^2$				
11/16/2016 13:00	66.5	70.0	-5.0	-4.964	25.000	5.000	25.000				
11/17/2016 13:00	69.0	70.0	-1.4	75th Percentile	2.041	1.429	2.041	n	S_d	S_{d2}	$\sum\|d\|$
11/18/2016 13:00	68.3	70.0	-2.4	-2.571	5.898	2.429	5.898	14	1.430	10.137	51.571
11/19/2016 13:00	66.3	70.0	-5.3		27.939	5.286	27.939	n-1	$\sum d$	$\sum d^2$	$\sum\|d\|^2$
11/20/2016 13:00	66.4	70.0	-5.1		26.449	5.143	26.449	13	-51.571	216.551	216.551
11/21/2016 13:00	66.1	70.0	-5.6		31.041	5.571	31.041				
11/22/2016 13:00	66.6	70.0	-4.9		23.592	4.857	23.592				
11/23/2016 13:00	67.4	70.0	-3.7		13.796	3.714	13.796				
11/24/2016 13:00	67.0	70.0	-4.3		18.367	4.286	18.367				
11/25/2016 13:00	67.9	70.0	-3.0		9.000	3.000	9.000				
11/26/2016 13:00	67.1	70.0	-4.1		17.163	4.143	17.163				
11/27/2016 13:00	68.7	70.0	-1.9		3.449	1.857	3.449				
11/28/2016 13:00	67.8	70.0	-3.1		9.878	3.143	9.878				
11/29/2016 13:00	68.8	70.0	-1.7		2.939	1.714	2.939				

"AB" (Eqn 4)	
	3.684
"AS" (Eqn 5)	
	1.430

Bias (%) (Eqn 3)	
	4.36
Signed Bias (%)	
-4.36	

Both Signs Positive
FALSE
Both Signs Negative
TRUE

CV (%) (Eqn 2)
1.94

Upper Probability Limit	Lower Probability Limit
-0.88	-6.49



Meteorological Summary




Pressure(hPa)
1040.0
1015.0
990.0
RH(%)
200.0
0.0
SR(w/m^2)
2000.0
0.0
Precip(mm)
10.0
0.0
16 17 18 19 20 21 22 23 24 25 26 27 28 29